

# Unsatisfactory rates vary between cervical cytology samples prepared using ThinPrep and SurePath platforms: a review and meta-analysis

Daniel Fontaine,<sup>1</sup> Nadira Narine,<sup>2</sup> Christopher Naugler<sup>1</sup>

**To cite:** Fontaine D, Narine N, Naugler C. Unsatisfactory rates vary between cervical cytology samples prepared using ThinPrep and SurePath platforms: a review and meta-analysis. *BMJ Open* 2012;**2**:e000847. doi: 10.1136/bmjopen-2012-000847

► Prepublication history for this paper is available online. To view this file please visit the journal online (<http://dx.doi.org/10.1136/bmjopen-2012-000847>).

Received 8 January 2012  
Accepted 12 March 2012

This final article is available for use under the terms of the Creative Commons Attribution Non-Commercial 2.0 Licence; see <http://bmjopen.bmj.com>

<sup>1</sup>Department of Pathology and Laboratory Medicine, University of Calgary and Calgary Laboratory Services, Calgary, Alberta, Canada

<sup>2</sup>Cytology Department, Central Manchester Hospitals NHS Foundation Trust, Manchester, UK

## Correspondence to

Dr Christopher Naugler;  
[christopher.naugler@cls.ab.ca](mailto:christopher.naugler@cls.ab.ca)

## ABSTRACT

**Objective:** To compare unsatisfactory rates between the two major liquid-based cytology (LBC) platforms, namely ThinPrep (Hologic) and SurePath (Becton Dickinson).

**Design:** The authors performed both a systematic review and a meta-analysis. Inclusion criteria were English language, data presented on unsatisfactory rates for either ThinPrep or SurePath, utilising actual patient samples (ie, not laboratory manipulated samples) and no manipulation using acetic acid to increase the satisfactory rate. The authors searched PubMed for articles using the keywords 'SurePath' or 'ThinPrep' and 'unsatisfactory'. References of retrieved studies were searched for additional articles. Key researchers in the field were also contacted.

**Participants and interventions:** Eligible studies were reviewed for rates of unsatisfactory cervical cytology smears processed on either the ThinPrep or SurePath platforms (compared with a general linear model) or data on unsatisfactory rates for both platforms for the same laboratory and the same patient population (compared with a meta-analysis using a random effects model and pooled RR).

**Primary Outcome Measure:** Unsatisfactory rate of cervical cytology smears.

**Results:** A total of 1 120 418 cervical cytology smears were reported in 14 different studies using the SurePath platform for an overall unsatisfactory rate (weighted average) of 0.3%. 28 studies reported on 1 148 755 smears prepared using the ThinPrep platform for an overall unsatisfactory rate (weighted average) of 1.3%. The general linear model did not show a difference between LBC platforms when other variables were controlled for; however, the power to detect a difference (0.087) was very low. The meta-analysis performed on four studies where both ThinPrep and SurePath results were reported from the same laboratory showed fewer unsatisfactory tests from the SurePath platform (RR 0.44, 95% CI 0.25 to 0.77,  $p=0.004$ ).

**Conclusions:** Multiple factors affect LBC unsatisfactory rates. In a meta-analysis, cervical cytology samples prepared on the SurePath platform show significantly fewer unsatisfactory smears than those prepared on the ThinPrep platform.

## ARTICLE SUMMARY

### Article focus

- Two major LBC platforms are commonly used (ThinPrep (Hologic) and SurePath (Becton Dickinson))
- These employ different sample preparation methodologies, which may result in different unsatisfactory rates.

### Key messages

- Multiple factors are associated with LBC unsatisfactory rates.
- Cervical cytology samples prepared on the SurePath platform show significantly fewer unsatisfactory smears than those prepared on the ThinPrep platform.
- Unsatisfactory samples represent a missed opportunity for screening

### Strengths and limitations of this study

- Large sample size from multiple studies
- Relatively few studies performed a head-to-head evaluation of the two platforms
- Restriction to English literature and use of only one bibliographic search base (Medline) are limitations.

## INTRODUCTION

Since first introduced in the late 1990s, liquid-based cytology (LBC) was quickly adopted by many laboratories as a great improvement in the performance of gynaecologic cytology. The method permits laboratories to create slides rather than having prepared slides sent to them in various degrees of fixation and preparation. The LBC methodology also provides a cleaner smear as the proprietary methods remove obscuring elements such as blood and inflammation. This cleaning of samples is often cited as the reason there are lower unsatisfactory rates in LBC.<sup>1</sup> The liquid

## Liquid-based cytology unsatisfactory rates

platforms also allow for re-ex testing for human papilloma virus for borderline cases. However, large trials and a meta-analysis have recently shown that the performance of LBC may be equivalent to conventional cytology, thus making it difficult for laboratories to justify the added expense of LBC.<sup>2-4</sup>

When considering LBC, laboratories are faced with a decision between two main platforms produced by ThinPrep (Hologic) and SurePath (Becton Dickinson). Both technologies are Food and Drug Administration approved but employ different methodologies in sample preparation. In light of these differences in preparation, we set out to evaluate the existing English literature comparing the performance of each platform with respect to unsatisfactory rates.

### METHODS

#### Information sources and search strategy

Our meta-analysis adhered to the PRISMA statement for reporting on meta-analyses.<sup>5</sup> The initial literature search for articles was conducted using PubMed. Articles published in English between 1 January 1990 and 1 August 2011 were retrieved using the search words unsatisfactory and ThinPrep or SurePath. These date range were chosen to coincide with the period following the introduction of The Bethesda System (TBS) in 1989.<sup>6</sup> The references of the retrieved papers were examined for further possible studies not detected in the initial literature search. Additionally, we searched major clinical trial registries (<http://clinicaltrials.gov>, <http://isrctn.org>, <http://www.umin.ac.jp/ctr/index/htm/>, <http://www.anzctr.org.au/Default.aspx>, <http://www.trialregister.nl/trialreg/index.asp>, <http://apps.who.int/trialsearch>) using the search words liquid-based cytology, contacted both manufacturers and contacted key researchers in the field to check for ongoing or unpublished studies.

#### Study selection and eligibility criteria

Eligible studies for the review presented rates of technically unsatisfactory cervical cytology smears performed on either ThinPrep or SurePath platform. For inclusion in the meta-analysis, studies must have presented data on unsatisfactory cervical cytology smears performed on both ThinPrep and SurePath platforms by the same laboratory on the same patient population. Only studies presented in English were considered. Both published and unpublished studies were considered to be eligible for inclusion.

The initial search retrieved 76 papers. The abstracts of these papers were reviewed by CN and DF and 46 were chosen for full review. An additional six published studies were subsequently identified as well as one unpublished study. Of these articles, 42 met the further inclusion criteria of presenting data on unsatisfactory rates for either ThinPrep or SurePath and utilising actual patient samples (ie, not laboratory manipulated samples) and not employing glacial acetic acid to


increase the satisfactory rate. Four of these studies presented data on both ThinPrep and SurePath use in the same population by the same laboratory and were therefore included in the meta-analysis. Our contacts with industry and other researchers did not reveal additional usable studies. The selection process is summarised in [figure 1](#).

#### Data extraction

Two of the authors (CN and DF) extracted data independently from the 42 studies meeting the inclusion criteria. These results were then compared and any discrepancies resolved. The following data were extracted: (1) number of patients included, (2) type of liquid-based system used (ThinPrep or SurePath), (3) per cent of unsatisfactory cases, (4) country of study, (5) year of study, (6) whether image analysis software was used to screen the slides and (7) the number of women in the study (sample size). The lack of ambiguity of the outcome measure (only two end points are possible: unsatisfactory or satisfactory), combined with the irrelevance of a follow-up (latent) period, obviated the need for quality scoring of the included studies.

#### Statistical analysis

Unsatisfactory rates were first compared using a univariate general linear model in SPSS V.19. The dependent variable was unsatisfactory rate, and the dependent variables were year of study (divided into 2002 and before to capture studies published before TBS revision in 2001 and 2003 and later), platform (ThinPrep vs SurePath), country where study was performed and whether image analysis software was used. Studies were weighted by the number of women participating in each one. In addition to the significance level, observed power was calculated for each variable. Differences were considered significant at an  $\alpha$  level of 0.05. Meta-analysis was then performed using Review Manager V.4.2.<sup>7</sup> A  $\chi^2$  test for heterogeneity among studies included in the meta-analysis revealed an  $I^2$  of 97% with a  $p$  value of  $<0.00001$ , indicating significant



**Figure 1** Selection of studies of unsatisfactory rates of cervical cytology smears prepared with ThinPrep and SurePath platforms.

heterogeneity among these studies (although the heterogeneity was quantitative rather than qualitative as all the trials demonstrated the same trend). Therefore, a random effects model was employed to calculate the pooled OR.

## RESULTS

A total of 1 120 418 cervical cytology smears were reported in 14 different studies using the SurePath platform (table 1). The pooled unsatisfactory rate (weighted average) was 0.3%. For cervical cytology smears processed on the ThinPrep platform, 28 studies reported on 1 148 755 smears for a pooled unsatisfactory rate (weighted average) of 1.3% (table 2). A general linear model as previously described was employed to test for a difference in unsatisfactory rate ThinPrep and SurePath. Year of publication ( $p=0.021$ ) and country where research was performed ( $p=0.004$ ) were independent predictors of unsatisfactory rate. The use of image analysis software ( $p=0.091$ ) and the LBP platform used ( $p=0.558$ ) were not independent predictors of unsatisfactory rate. However, the observed power for LBP platform was very low at 0.087, making the risk of a type II error high. To attempt to gain statistical power, we performed a meta-analysis where we considered only studies where both platforms were evaluated on the same patient population by the same laboratory. We found three such studies in the published literature and added data from one additional unpublished study (table 3). All four of these studies reported similar populations but different patients who were screened with either ThinPrep or SurePath platform. The pooled RR from the meta-analysis was 0.44 (95% CI 0.25 to

0.77) in favour of SurePath over ThinPrep (figure 2). A Z-test for overall effect was statistically significant at  $p=0.004$ .

Because of the small number of studies included in the meta-analysis, a funnel plot was not informative. However, the four included showed similar unsatisfactory rates to other studies and therefore we felt were representative of the other studies retrieved.

## DISCUSSION

Despite being one of the most successful screening tests for cancer, cervical cytology smears have been criticised for low sensitivity. False-negative cervical cytology smears are responsible for this decrease in sensitivity and may result from collection errors, screening errors or interpretation errors.

We set out to examine unsatisfactory rates in LBC as these have been demonstrated to be largely reproducible within laboratories but not across laboratories.<sup>43</sup> Most laboratories use TBS or some modification thereof for interpretation of cervical cytology smears. We could not, however, establish the criteria used to determine adequacy for each platform; some laboratories use the methods described in the manufacturer's guide (TBS based), while others have adopted their own criteria and many authors do not elaborate on this in their methods. Furthermore, our results showed that year of publication and to a greater extent country were significant confounding factors in the comparison of unsatisfactory rates between ThinPrep and SurePath platforms. These confounding factors resulted in very low power to detect a statistically significant difference in unsatisfactory rates, despite a wide difference in mean unsatisfactory rates

**Table 1** Unsatisfactory cervical cytology smear rates in populations using the SurePath platform

Lead author	Year	Total number of cervical cytology smears	Number of unsatisfactory smears	Percentage of unsatisfactory smears	Location of study
Colgan <i>et al</i> <sup>8</sup>	2004	352 680	915	0.3	Canada
Zhao <i>et al</i> <sup>9</sup>	2011	972	2	0.2	China
Kirscher <i>et al</i> <sup>10</sup>	2006	84 414	292	0.3	Denmark
Beerman <i>et al</i> <sup>11</sup>	2008	35 315	46	0.1	the Netherlands
Sykes <i>et al</i> <sup>12</sup>	2008	451	12	2.7	New Zealand
Sykes <i>et al</i> <sup>12</sup>	2008	457	12	2.6	New Zealand
Kitchener <i>et al</i> <sup>13</sup>	2011	22 145	354	1.6	UK
Narine and Young <sup>14</sup>	2007	53 982	208	0.4	UK
Narine and Young <sup>14</sup>	Unpublished	27 738	327	1.2	UK
Alsharif <i>et al</i> <sup>15</sup>	2009	232 022	360	0.2	USA
Stark <sup>16</sup>	2007	137 703	302	0.2	USA
Nance <sup>17</sup>	2007	92 875	158	0.2	USA
Fremont-Smith <i>et al</i> <sup>18</sup>	2004	58 580	130	0.2	USA
Sass <sup>19</sup>	2003	8771	14	0.2	USA
Wilbur <i>et al</i> <sup>20</sup>	2009	12 313	27	0.2	USA
Total		1 120 418	3159	0.3	

## Liquid-based cytology unsatisfactory rates

**Table 2** Unsatisfactory cervical cytology smear rates in populations using the ThinPrep platform


Lead author	Year	Total number of cervical cytology smears	Number of unsatisfactory smears	Percentage of unsatisfactory smears	Location of study
Davey <i>et al</i> <sup>1</sup>	2007	52 665	795	1.5	Australia
Roberts <i>et al</i> <sup>22</sup>	1997	35 560	235	0.7	Australia
Shield <i>et al</i> <sup>23</sup>	1999	300	19	6.3	Australia
Halford <i>et al</i> <sup>24</sup>	2009	87 284	758	0.9	Australia
Duggan <i>et al</i> <sup>25</sup>	2006	2288	10	0.4	Canada
Belison <i>et al</i> <sup>26</sup>	2002	1450	158	10.9	China
Zhao <i>et al</i> <sup>9</sup>	2011	1033	15	1.5	China
Yeoh <i>et al</i> <sup>27</sup>	1999	16 541	93	0.6	Hong Kong
Cheung <i>et al</i> <sup>28</sup>	2003	190 667	608	0.3	Hong Kong
Grace <i>et al</i> <sup>29</sup>	2002	1000	13	1.3	Ireland
Treacy <i>et al</i> <sup>1</sup>	2009	41 312	1342	3.2	Ireland
Ronco <i>et al</i> <sup>30</sup>	2007	22 708	583	2.6	Italy
Rahimi <i>et al</i> <sup>31</sup>	2008	467	9	1.9	Italy
Luthra <i>et al</i> <sup>32</sup>	2001	1024	36	3.5	Kuwait
Siebers <i>et al</i> <sup>33</sup>	2008	46 064	153	0.3	the Netherlands
Williams <sup>34</sup>	2006	78 064	1493	1.9	Scotland
Taylor <i>et al</i> <sup>35</sup>	2006	3184	70	2.2	South Africa
Tuncer <i>et al</i> <sup>36</sup>	2005	4322	72	1.7	Turkey
Kitchener <i>et al</i> <sup>13</sup>	2011	23 436	456	1.9	UK
Narine and Young <sup>14</sup>	Unpublished	58 973	1250	2.1	UK
Alsharif <i>et al</i> <sup>15</sup>	2009	5419	124	2.3	USA
Nance <sup>17</sup>	2007	88 575	567	0.6	USA
Bentz <sup>37</sup>	2002	23 790	209	0.9	USA
Bentz <sup>37</sup>	2002	15 154	197	1.3	USA
Bolick and Hellman <sup>38</sup>	1998	10 694	31	0.3	USA
Wachtel <i>et al</i> <sup>39</sup>	2009	2890	77	2.7	USA
Harkness <i>et al</i> <sup>40</sup>	2003	3000	80	2.7	USA
Duby and DiFurio <sup>41</sup>	2009	55 438	1027	1.8	USA
Duby and DiFurio <sup>41</sup>	2009	53 209	783	1.5	USA
Zhao and Austin <sup>42</sup>	2009	222 242	3230	1.4	USA
Total		1 148 755	15 593	1.3	

**Table 3** Studies included in the meta-analysis

Lead author	Year	Platform	Total number of cervical cytology smears	Number of unsatisfactory smears	Percentage of unsatisfactory smears	Location of study
Zhao <i>et al</i> <sup>9</sup>	2011	TP	972	2	1.5	China
Zhao <i>et al</i> <sup>9</sup>	2011	SP	1033	15	0.2	China
Kitchener <i>et al</i> <sup>13</sup>	2011	TP	23 436	456	2.0	UK
Kitchener <i>et al</i> <sup>13</sup>	2011	SP	22 145	354	1.6	UK
Narine and Young <sup>14</sup>	Unpublished	TP	58 973	1250	2.1	UK
Narine and Young <sup>14</sup>	Unpublished	SP	27 738	327	1.2	UK
Nance <sup>17</sup>	2007	TP	88 575	567	0.6	USA
Nance <sup>17</sup>	2007	SP	92 875	158	0.2	USA

SP, SurePath; TP, ThinPrep.

**Figure 2** Meta-analysis of the association between rate of unsatisfactory cervical smears and processing platform (ThinPrep and SurePath).



between the two platforms. In contrast, a meta-analysis comparing four head-to-head studies did reveal a lower unsatisfactory rate for the SurePath platform. These differences underscore the fact that multiple factors are associated with overall unsatisfactory rates with the platform used representing only one of these.

It is also acknowledged that there are other possible biases in the source literature that our analyses could not control for. These include the relative experience of the cytologists in reading slides prepared using one or the other platform and any possible undisclosed biases in the allocation of individuals to treatment groups in the four head-to-head studies. The differences in unsatisfactory rates due to LBC platform we report here are most likely a function of differences in the proprietary methodology for each platform. ThinPrep employs a filter-based technology, while SurePath uses a density/sedimentation process. Differences are also encountered when collecting the sample; ThinPrep requires the collector to rinse the collection device in the liquid media that is followed by disposal of the collection device. In contrast, the SurePath collection device is placed in the media and sent to the laboratory for processing. This difference has been demonstrated to account for up to 38% loss of the ThinPrep sample in a study by Bigras *et al*.<sup>44</sup> These preparatory differences represent a significant technical difference, which laboratories must consider prior to adopting LBC. However, counterbalancing this loss of sample is the fact that ThinPrep samples are significantly less labour intensive than SurePath samples for the processing laboratory.

As tables 1 and 2 show, there was considerable variation in unsatisfactory rates among jurisdictions, with countries in the European Union in particular tending to show higher unsatisfactory rates in both studies of SurePath and ThinPrep. This is likely attributable to variations in adequacy criteria compared with TBS.

Both the SurePath (FocalPoint GS) and ThinPrep (ThinPrep Imager) systems include optional computer-assisted digital image analysis software. If we examine only studies using this technology, three studies using the ThinPrep Imager all gave unsatisfactory rates of 1.5%<sup>21 41 42</sup>—similar to other ThinPrep studies included in this review. The single study evaluating the FocalPoint GS system showed an unsatisfactory rate of 0.2%,<sup>19</sup> again similar to the other SurePath studies reviewed.

Reprocessing of unsatisfactory ThinPrep slides using glacial acetic acid can decrease the unsatisfactory rates by 30%–40%.<sup>45</sup> However, use of this technique still

would not result in equivalency with SurePath unsatisfactory rates and introduces additional issues including a higher false-positive rate<sup>46</sup> and interferes with the ability to perform hybrid capture human papilloma virus testing.<sup>47</sup>

The data presented here show a significant difference between the different platforms with SurePath having a significantly lower rate of unsatisfactory samples. This pattern was observed when considering different countries and methods used to determine thresholds for unsatisfactory designation. Consistent with this observation are data from a College of American Pathologist survey reporting differences in unsatisfactory rates between laboratories using SurePath (median unsatisfactory rate 0.3%) and ThinPrep (median unsatisfactory rate 1.1%) platforms.<sup>48</sup> These numbers are similar to the pooled unsatisfactory rates we report in this study (0.3% for SurePath and 1.3% for ThinPrep).

A limitation of our study is that it considers only one aspect of LBP platform selection (unsatisfactory rate), whereas other factors such as specificity and sensitivity must also be considered. A discussion of this is outside the scope of this paper but interested readers are referred to the meta-analysis of Arbyn *et al*<sup>4</sup> for further information. The implication of this unsatisfactory rate difference is also a consideration for specimens, which could be biobanked for further studies to evaluate new biomarkers and how they may perform in unsatisfactory LBC samples.<sup>49</sup>

Unsatisfactory samples represent a missed opportunity for screening<sup>50</sup> and are more often associated with a cervical abnormality.<sup>51</sup> It has also been well described that patients with an invasive cervical cancer have a much greater rate of unsatisfactory Pap smears, often as a result of scant cellularity and obscuring elements such as blood and inflammation.<sup>52–54</sup> In addition to the technical requirements of sample preparation, laboratories may wish to consider the variation in unsatisfactory sample rates when choosing an LBC platform.

**Contributors** DF, NN and CN designed the study. DF and CN collected the data. CN analysed and interpreted the data. The manuscript was prepared by DF and CN and commented on by all authors. CN is the guarantor.

**Funding** This research received no specific grant from any funding agency in the public, commercial or not-for-profit sectors.

**Competing interests** None.

**Provenance and peer review** Not commissioned; externally peer reviewed.

**Data sharing statement** Data used in this paper are available from the corresponding author.

## Liquid-based cytology unsatisfactory rates

### REFERENCES

- Treacy A, Reynolds J, Kay EW, *et al*. Has the ThinPrep method of cervical screening maintained its improvement over conventional smears in terms of specimen adequacy? *Diagn Cytopathol* 2009;37:239–40.
- Confortini M, Bergeron C, Desai M, *et al*. Accuracy of liquid-based cytology: comparison of the results obtained within a randomized controlled trial (the New Technologies for Cervical Cancer Screening Study) and an external group of experts. *Cancer Cytopathol* 2010;118:203–8.
- Castle PE, Bulten J, Confortini M, *et al*. Age-specific patterns of unsatisfactory results for conventional Pap smears and liquid-based cytology: data from two randomised clinical trials. *BJOG* 2010;117:1067–73.
- Arbyn M, Bergeron C, Klinkhamer P, *et al*. Liquid compared with conventional cervical cytology: a systematic review and meta-analysis. *Obstet Gynecol* 2008;111:167–77.
- Moher D, Liberati A, Tetzlaff J, *et al*. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *Ann Intern Med* 2009;151:264–9, W64.
- National Cancer Institute Workshop. The 1988 Bethesda System for reporting cervical/vaginal cytological diagnoses. *JAMA* 1989;262:931–4.
- Review Manager (RevMan) version 4.2 for Windows*. Copenhagen: The Nordic Cochrane Centre, The Cochrane Collaboration, 2003.
- Colgan TJ, McLachlin CM, Cotterchio M, *et al*. Results of the implementation of liquid-based cytology—SurePath in the Ontario screening program. *Cancer* 2004;102:362–7.
- Zhao FH, Hu SY, Bian JJ, *et al*. Comparison of ThinPrep and SurePath liquid-based cytology and subsequent human papillomavirus DNA testing in China. *Cancer Cytopathol* 2011;119:387–94.
- Kirschner B, Simonsen K, Jungw J. Comparison of conventional Papanicolaou smear and SurePath liquid-based cytology in the Copenhagen population screening programme for cervical cancer. *Cytopathology* 2006;17:187–94.
- Beerman H, van Dorst EB, Kuenen-Boumeester V, *et al*. Superior performance of liquid-based versus conventional cytology in a population-based cervical cancer screening program. *Gynecol Oncol* 2009;112:572–6.
- Sykes PH, Harker DY, Miller A, *et al*. A randomized comparison of SurePath liquid-based cytology and conventional smear cytology in a colposcopy clinic setting. *BJOG* 2008;115:1375–81.
- Kitchener HC, Blanks R, Dunn G, *et al*. Automation-assisted versus manual reading of cervical cytology (MAVARIC): a randomised controlled trial. *Lancet Oncol* 2011;12:56–64.
- Narine N, Young W. Transformation zone sampling rate is a useful performance indicator for practitioners collecting cervical samples using SurePath liquid-based cytology system. *Cytopathology* 2007;18:220–4.
- Alsharif M, McKeon DM, Gulbahce HE. Unsatisfactory SurePath liquid-based Papanicolaou tests: causes and signi cance. *Cancer Cytopathol* 2009;117:15–26.
- Stark A. Rejected and unsatisfactory specimens: a comparative study of liquid-based (SurePath) preparations and conventional pap smears for cervicovaginal screening. *Lab Med* 2007;38:729–33.
- Nance KV. Evolution of pap testing at a community hospital: a ten year experience. *Diagn Cytopathol* 2007;35:248–53.
- Fremont-Smith M, Marino J, Griffin B, *et al*. Comparison of the Sure Path liquid-based Papanicolaou smear with the conventional Papanicolaou smear in a multi-site direct-to-vial study. *Cancer* 2004;102:269–79.
- Sass MA. Use of a liquid-based, thin-layer Pap test in a community hospital. Impact on cytology performance and productivity. *Acta Cytol* 2004;48:17–22.
- Wilbur DC, Black-Shaffer WS, Luff RD, *et al*. The Becton Dickinson FocalPoint GS Imaging System: clinical trials demonstrate significantly improved sensitivity for the detection of important cervical lesions. *Am J Clin Pathol* 2009;132:767–75.
- Davey E, d Assuncao J, Irwig L, *et al*. Accuracy of reading liquid based cytology slides using the ThinPrep Imager compared with conventional cytology: prospective study. *BMJ* 2007;335:31.
- Roberts JM, Gurley AM, Thurloe JK, *et al*. Evaluation of the ThinPrep Pap test as an adjunct to the conventional Pap smear. *Med J Aust* 1997;167:466–9.
- Shield PW, Nolan GR, Phillips GE, *et al*. Improving cervical cytology screening in a remote, high risk population. *Med J Aust* 1999;170:255–8.
- Halford JA, Batty T, Boost T, *et al*. Comparison of the sensitivity of conventional cytology and the ThinPrep Imaging System for 1,083 biopsy confirmed high-grade squamous lesions. *Diagn Cytopathol* 2010;38:318–26.
- Duggan MA, Khalil M, Brasher PM, *et al*. Comparative study of the ThinPrep Pap test and conventional cytology results in a Canadian cohort. *Cytopathology* 2006;17:73–81.
- Belinson JL, Pan QJ, Biscotti C, *et al*. Primary screening with liquid-based cytology in an unscreened population in rural China, with an emphasis on reprocessing unsatisfactory samples. *Acta Cytol* 2002;46:470–4.
- Yeoh GP, Chan KW, Lauder I, *et al*. Evaluation of the ThinPrep Papanicolaou test in clinical practice 6-month study of 16,541 cases with histological correlation in 220 cases. *Hong Kong Med J* 1999;5:233–9.
- Cheung AN, Szeto EF, Leung BS, *et al*. Liquid-based cytology and conventional cervical smears: a comparison study in an Asian screening population. *Cancer* 2003;99:331–5.
- Grace A, McBrearty P, Troost S, *et al*. Comparative study: conventional cervical and ThinPrep Pap tests done in a routine clinical setting. *Cytopathology* 2002;13:200–5.
- Ronco G, Cuzick J, Pierotti P, *et al*. Accuracy of liquid based versus conventional cytology: overall results of new technologies for cervical cancer screening: randomised controlled trial. *BMJ* 2007;335:28.
- Rahimi S, Carnovale-Scalzo C, Marani C, *et al*. Comparison of conventional Papanicolaou smears and liquid-based, thin-layer cytology with colposcopic biopsy control in central Italy: a consecutive sampling study of 461 cases. *Diagn Cytopathol* 2009;37:1–3.
- Luthra UK, Chishti M, Dey P, *et al*. Performance of monolayered cervical smears in a gynecology outpatient setting in Kuwait. *Acta Cytol* 2002;46:303–10.
- Siebers AG, Klinkhamer PJ, Arbyn M, *et al*. Cytologic detection of cervical abnormalities using liquid-based compared with conventional cytology: a randomized controlled trial. *Obstet Gynecol* 2008;112:1327–43.
- Williams AR. Liquid-based cytology and conventional smears compared over two 12-month periods. *Cytopathology* 2006;17:82–5.
- Taylor S, Kuhn L, Dupree W, *et al*. Direct comparison of liquid-based and conventional cytology in a South African screening trial. *Int J Cancer* 2006;118:957–62.
- Tuncer ZS, Basaran M, Sezgin Y. Clinical results of a split sample liquid-based cytology (ThinPrep) study of 4,322 patients in a Turkish institution. *Eur J Gynaecol Oncol* 2005;26:646–8.
- Bentz JS. Liquid-based cytology for cervical cancer screening. *Expert Rev Mol Diagn* 2005;5:857–71.
- Bolick DR, Hellman DJ. Laboratory implementation and efficacy assessment of the ThinPrep cervical cancer screening system. *Acta Cytol* 1998;42:209–13.
- Wachtel MS, Hatley WG, de Riese C. Using poisson regression to compare rates of unsatisfactory pap smears among gynaecologists and to evaluate a performance improvement plan. *Acta Cytol* 2009;53:160–4.
- Harkness CB, Theofrastous JP, Ibrahim SN, *et al*. Papanicolaou and thin-layer cervical cytology with colposcopic biopsy control. A comparison. *J Reprod Med* 2003;48:681–6.
- Duby JM, DiFurio MJ. Implementation of the ThinPrep Imaging System in a tertiary military medical centre. *Cancer Cytopathol* 2009;117:264–70.
- Zhao C, Austin RM. High-risk human papillomavirus DNA test results are useful for disease risk stratification in women with unsatisfactory liquid-based cytology pap test results. *J Low Genit Tract Dis* 2009;13:79–84.
- Haroon S, Samayoa L, Witzke D, *et al*. Reproducibility of cervicovaginal ThinPrep cellularity assessment. *Diagn Cytopathol* 2002;26:19–21.
- Bigras G, Rieder MA, Lamercy JM, *et al*. Keeping collecting device in liquid medium is mandatory to ensure optimized liquid-based cervical cytologic sampling. *J Low Genit Tract Dis* 2003;7:168–74.
- Islam S, West AM, Saboorian MH, *et al*. Reprocessing unsatisfactory ThinPrep Papanicolaou test specimens increases sample adequacy and detection of significant cervicovaginal lesions. *Cancer* 2004;102:67–73.
- Rowe LR, Bentz JS. A simple method to determine the need for glacial acetic acid treatment of bloody ThinPrep Pap tests before slide processing. *Diagn Cytopathol* 2004;31:321–5.
- Agoff SN, Dean T, Nixon BK, *et al*. The efficacy of reprocessing unsatisfactory cervicovaginal ThinPrep specimens with and without glacial acetic acid: effect on Hybrid Capture II human papillomavirus testing and clinical follow-up. *Am J Clin Pathol* 2002;118:727–32.
- Moriarty AT, Clayton AC, Zaleski S, *et al*. Unsatisfactory reporting rates: 2006 practices of participants in the college of american pathologists interlaboratory comparison program in gynecologic cytology. *Arch Pathol Lab Med* 2009;133:1912–16.

## Liquid based cytology unsatisfactory rates

49. Arbyn M, Van Veen EB, Andersson K, *et al.* Cervical cytology biobanking in Europe. *Int J Biol Markers* 2010; 25:117–25.
50. Bentz JS, Rowe LR, Gopez EV, *et al.* The unsatisfactory ThinPrep Pap test: missed opportunity for disease detection? *Am J Clin Pathol* 2002;117:457–63.
51. Nance KV. Unsatisfactory cervical cytology screening specimens are associated with missed cervical dysplasia and cancer. *BJOG* 2009;116:866–7.
52. Ransdell JS, Davey DD, Zaleski S. Clinicopathologic correlation of the unsatisfactory Papanicolaou smear. *Cancer* 1997;81:139–43.
53. Paterson ME, Peel KR, Joslin CA. Cervical smear histories of 500 women with invasive cervical cancer in Yorkshire. *Br Med J (Clin Res Ed)* 1984;289:896–8.
54. Sherman ME, Kelly D. High-grade squamous intraepithelial lesions and invasive carcinoma following the report of three negative Papanicolaou smears: screening failures or rapid progression? *Mod Pathol* 1992;5:337–42.



# Unsatisfactory rates vary between cervical cytology samples prepared using ThinPrep and SurePath platforms: a review and meta-analysis

Daniel Fontaine, Nadira Narine and Christopher Naugler

*BMJ Open* 2012 2:

doi: 10.1136/bmjopen-2012-000847

---

Updated information and services can be found at:

<http://bmjopen.bmj.com/content/2/2/e000847.full.html>

---

*These include:*

**Data Supplement**

"Supplementary Data"

<http://bmjopen.bmj.com/content/suppl/2012/04/21/bmjopen-2012-000847.DC1.html>

**References**

This article cites 53 articles, 6 of which can be accessed free at:

<http://bmjopen.bmj.com/content/2/2/e000847.full.html#ref-list-1>

**Open Access**

This is an open-access article distributed under the terms of the Creative Commons Attribution Non-commercial License, which permits use, distribution, and reproduction in any medium, provided the original work is properly cited, the use is non commercial and is otherwise in compliance with the license. See:

<http://creativecommons.org/licenses/by-nc/2.0/> and

<http://creativecommons.org/licenses/by-nc/2.0/legalcode>.

**Email alerting service**

Receive free email alerts when new articles cite this article. Sign up in the box at the top right corner of the online article.

---

**Topic Collections**

Articles on similar topics can be found in the following collections

[Diagnostics](#) (18 articles)

[Pathology](#) (8 articles)

[Sexual health](#) (20 articles)

---

To request permissions go to:

<http://group.bmj.com/group/rights-licensing/permissions>

To order reprints go to:

<http://journals.bmj.com/cgi/reprintform>

To subscribe to BMJ go to:

<http://group.bmj.com/subscribe/>

## Notes

---

To request permissions go to:

<http://group.bmj.com/group/rights-licensing/permissions>

To order reprints go to:

<http://journals.bmj.com/cgi/reprintform>

To subscribe to BMJ go to:

<http://group.bmj.com/subscribe/>